



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**

Customer PO # :

**Packaged Product:** THE294  
 CONDUCTIVITY STANDARD 575 UMHOS @25 C (1L)

**Lot #** 1041455

**Chemical Product:** THE294F  
 CONDUCTIVITY STANDARD 575 UMHOS @25 C

Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES YES		YES
Sample ID			
Tested 04/15/21			
Tester DTRA			
Recorded 04/15/21			
QCSP30: CONDUCTIVITY STD&MEASUREMENT	YES QCSP30		YES
Sample ID			
Tested 04/15/21			
Tester DTRA			
Recorded 04/15/21			
TRACEABLE TO N.I.S.T. SRM 3191	YES YES		YES
Sample ID			
Tested 04/15/21			
Tester DTRA			
Recorded 04/15/21			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 04/15/21			
Tester DTRA			
Recorded 04/15/21			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M92,T234
Sample ID			
Tested 04/15/21			
Tester DTRA			
Recorded 04/15/21			
CONDUCTIVITY (US/CM)	575 US/CM	573 - 577	575
Sample ID			
Tested 04/15/21			
Tester DTRA			
Recorded 04/15/21			

**Made** 04/15/21