



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536  
 Website: [www.aquasolutions.org](http://www.aquasolutions.org)

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
 Customer PO # :

<b>Packaged Product:</b> 3858-1L FLUORIDE STANDARD 1.0 PPM (1L)			
Lot # 1042886			
<b>Chemical Product:</b> 3858 FLUORIDE STANDARD 1.0 PPM			
<b>Test</b>	<b>Target/UOM</b>	<b>Range</b>	<b>Result</b>
PREPARED TO FORMULATION	YES YES		YES
Sample ID			
Tested 04/28/21			
Tester LLAC			
Recorded 04/28/21			
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
Sample ID			
Tested 04/28/21			
Tester LLAC			
Recorded 04/28/21			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		P35,VF3-J10,M79
Sample ID			
Tested 04/28/21			
Tester LLAC			
Recorded 04/28/21			
PPM FLUORIDE	1.00 PPM	0.85 - 1.15	1.05
Sample ID			
Tested 04/28/21			
Tester LLAC			
Recorded 04/28/21			
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		10/30/21
Sample ID			
Tested 04/28/21			
Tester LLAC			
Recorded 04/28/21			
SHELF LIFE OF PRODUCT	6 MONTHS TIME		6 MONTHS
Sample ID			
Tested 04/28/21			
Tester LLAC			
Recorded 04/28/21			
Made	04/28/21		



Aqua Solutions, Inc.  
Ph: 281-479-2569  
Fax: 281-479-2790  
P.O. Box 70  
Deer Park, TX 77536

Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 2

Date: 04/30/21 at 9:06 AM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**