

CORNING

Glass Certification

Corning Pharmaceutical Glass, LLC
 563 Crystal Ave
 Vineland, NJ 08360 US
 (856) 794-7100

TO:
 ACME VIAL & GLASS CO
 1601 COMMERCE WAY
 PASO ROBLES, CA 93446 US
 805-239-2666
 lisa@acmevial.com

Delivery No. / Date
 83016433 / 11/24/2021

Item	Article	Pallet ID	Sales Ord. Pal Qty	Cust. PO # Batch	Quantity ProdOrd	UOM
010	7014753003 TPO401475R06973 C51-L, 14.75MMOD X 1.00MMWW X 1500MM GMA Corning® 51-L AMBER		3025483	4875	11,930.0	LB
	PALLET ID:					
		1107210853311	814.0 LB	NOV2101181	36051467	
		1107211041011	818.0 LB	NOV2101183	36051467	
		1107211147511	809.0 LB	NOV2101196	36051467	
		1107211310011	806.0 LB	NOV2101208	36051467	
		1107211434511	795.0 LB	NOV2101219	36051467	
		1107211552411	814.0 LB	NOV2101227	36051467	
		1107211740411	803.0 LB	NOV2101236	36051467	
		1107211840411	785.0 LB	NOV2101251	36051467	
		1107212004511	790.0 LB	NOV2101259	36051467	
		1107212128211	809.0 LB	NOV2101275	36051467	
		1107212248511	818.0 LB	NOV2101281	36051467	
		1107212353511	821.0 LB	NOV2101289	36051467	
		1108210353311	703.0 LB	NOV2101319	36051467	
		1108212309311	770.0 LB	NOV2101479	36051467	
		1109210036111	775.0 LB	NOV2101483	36051467	
020	9009801002 C51-V, 9.80MMOD X 0.60MMWW X 1550MM Corning® 51-V CLEAR		3025982	4933	1,335.0	LB
	PALLET ID:					
		1115212253520	1,335.0 LB	NOV2102894	36051921	
030	9018751010 C51-V, 18.75MMOD X 1.15MMWW X 1524MM Corning® 51-V CLEAR		3025828	4913	7,209.0	LB
	PALLET ID:					
		1030211543018	2,403.0 LB	OCT2104764	36052148	
		1030211733418	2,403.0 LB	OCT2104775	36052148	
		1030211937418	2,403.0 LB	OCT2104787	36052148	
040	10003 0006020 00000 PALLET, CHR, HT, 48X40X5.56, TUBE, SPEC				19	EA

Certificate of Conformity

Glass Tubing manufactured in
 Vineland, New Jersey in the United States of America

Corning® 33 Type I CLEAR Borosilicate Tubing, Corning® 51-V Type I CLEAR Borosilicate Tubing, and Corning® 51-L Type I AMBER Borosilicate Tubing are produced in accordance with ISO 9001:2015 and ISO 15378:2015 quality Standards with the following conformance statements:

I. USP/EP/JP - TYPE I GLASS GRAINS

Corning clear (flint) and amber tubing is made of Type I Borosilicate glass and meets the requirements of the glass grains test according to the most current editions of the United States Pharmacopeia or USP <660>, European Pharmacopoeia or EP (3.2.1 B), and any other Pharmacopoeias, which are in accordance with USP or EP definitions of Type I Glass as per glass grains method. The glass conforms to JP 7.01 Test for Glass Containers for injections, (3) Soluble Alkali test, Method 1.