



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer: DEP040 / SCI (SCIENTIFIC CONSUMABLES&INSTRU)

Packaged Product:	6600-500ML PERCHLORIC ACID TITRANT 0.1 NORMAL (500ML)		
Lot #	1101141		
Chemical Product:	6600 PERCHLORIC ACID TITRANT 0.1 NORMAL ACETOUS		
Test	Target/UOM	Range	Result
Actual Normality	0.1000 EQ/L	0.0995 - 0.1005	0.0999
QCSP16:STANDARDIZATION OF ACETOUS PERCHL	YES QCSP16		YES
N.I.S.T. Traceable to SRM 84	Yes YES		YES
PREPARED TO FORMULATION ON FILE	YES YES		YES
APPEARANCE AND COLOR	CLEAR WATER WHITE LIQUID APPEARANCE		PASS
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		(O)T-15,M83,EF6-XXX,M51,BID 02185
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		10/30/22
%WATER	0.020-0.050 %WATER		0.030
Actual Density	1.05793 GM/ML	1.05761 - 1.05825	1.05804
AMENDMENT TO LOT#:	LOT # HAS BEEN REVISED LOT NUMBER		1101141
Made	10/18/21		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager