

Certificate of Analysis

Doctor Solution (Sodium Plumbite)

Lot Number: 1101A44

Product Number: 2650

Manufacture Date: JAN 19, 2021

Expiration Date: JAN 2023

| Name | CAS# | Grade |
|------------------|-----------|-----------------|
| Water | 7732-18-5 | ACS/ASTM/USP/EP |
| Sodium Hydroxide | 1310-73-2 | ACS |
| Lead Monoxide | 1317-36-8 | ACS |

| Test | Specification | Result |
|------------|--|--------|
| Appearance | Colorless liquid with slight orange sediment | Passed |

| Specification | Reference |
|---------------|-----------|
|---------------|-----------|

Doctor (Sodium Plumbite) Solution

ASTM (D 4952)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 2650-16 | 500 mL natural poly | 24 months |
| 2650-32 | 1 L natural poly | 24 months |

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Israel Alamudun (01/19/2021)

Quality Control Supervisor

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.