



Aqua Solutions, Inc.  
 Ph: 281-479-2569  
 Fax: 281-479-2790  
 P.O. Box 70  
 Deer Park, TX 77536

Website: [www.aquasolutions.org](http://www.aquasolutions.org)

Page: 1

Date: 12/07/21 at 1:06 PM

**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**

**Certificate of Analysis / QC Results**

Customer: DEP040 / SCI (SCIENTIFIC CONSUMABLES&INSTRU)

**Packaged Product:** 0295-4L  
 ALKALINE TITRATION SOLV.-ASTM D3227-13 (4L)

**Lot #** 1102802

**Chemical Product:** 0295  
 ALKALINE TITRATION SOLV.-ASTM D3227-13

Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Product nitrogen purged	YES		YES
IPA TESTED FOR PEROXIDES	PASSES		PASSES
APPEARANCE AND COLOR	CLEAR LIQUID		PASS
INSTRUMENTS USED DURING PREPARATION	APPEARANCE INSTRUMENT		VF12-83,M103,T-156,
EXPIRATION DATE	MM/DD/YY		11/30/22
SHELF LIFE OF PRODUCT	SHELF LIFE TIME		1 YEAR

**Made** 11/04/21

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

**Quality Manager**