RICCA CHEMICAL COMPANY®

Certificate of Analysis

Buffer, Reference Standard, pH 5.00 ± 0.01 at 25° C

Hou runnou rroocoo	\mathbf{Lot}	Number:	1105903
--------------------	----------------	---------	---------

Product Number: 1505

Manufacture Date: MAY 12, 2021 Expiration Date: APR 2023

The certified value for this product is confirmed in independent testing by a second qualified chemist. The NIST Traceable pH value is certified to ± 0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05 .

°C pH	$\begin{array}{c} 10 \\ 4.96 \end{array}$	$\begin{array}{c} 15 \\ 4.97 \end{array}$	$\begin{array}{c} 20 \\ 4.98 \end{array}$	$\begin{array}{c} 25 \\ 5.00 \end{array}$	$\begin{array}{c} 30 \\ 5.01 \end{array}$	$\begin{array}{c} 35 \\ 5.03 \end{array}$	$\begin{array}{c} 40 \\ 5.04 \end{array}$	$\begin{array}{c} 45 \\ 5.06 \end{array}$	$\begin{array}{c} 50 \\ 5.08 \end{array}$		
Name					CAS#				Grade		
Water					7732-18-5				ACS/ASTM/USP/EP		
Potass	ium Ac	id Phtha	late			87	7-24-7			Buffer	
Sodium Hydroxide				1310-73-2				Reagent			
Preservative					Proprietary						
Test					Specification				Result	NIST SRM#	
Appearance				Colorless liquid				Passed			
pH at 25°C (Method: SQCP027, SQCP033)				4.990-5.010				5.000	185i, 186-I-g, 186-II-g		
pH at 25°C (Method: SQCP027, SQCP033)					0.01				0.01		

Uncertainty

pH measurements were performed in our Arlington, TX laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution.

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1505-1	4 L natural poly	24 months
1505-16	500 mL natural poly	24 months
1505-5	20 L Cubitainer®	24 months
D 1104 1500		

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Israel Alamudun (05/12/2021) Quality Control Supervisor This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."



This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.