



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Test	Target/UOM	Range	Result
Packaged Product: 1500			
	BUFFER REFERENCE pH 7.0/YELLOW		
PREPARED TO FORMULATION ON	YES		YES
FILE	YES		
Sample ID			
Tested 11/04/21			
Tester RSWE			
Recorded 11/04/21			
PH OF PRODUCT/SOLUTION	7.000	6.990 - 7.010	6.991
	PH		
Sample ID			
Tested 11/04/21			
Tester RSWE			
Recorded 11/04/21			
QCSP 20 PH CALIBRATION	YES		YES
METHODS	METHOD		
Sample ID			
Tested 11/04/21			
Tester RSWE			
Recorded 11/04/21			
Actual Temp	25.0	24.9 - 25.1	25.0
	DEG C		
Sample ID			
Tested 11/04/21			
Tester RSWE			
Recorded 11/04/21			
pH Traceable to N.I.S.T. SRM'S	YES		YES
	YES		
Sample ID			
Tested 11/04/21			
Tester RSWE			
Recorded 11/04/21			
PH NIST TRACEABLE TO SRM 185	YES		YES
	YES		
Sample ID			
Tested 11/04/21			
Tester RSWE			
Recorded 11/04/21			
NIST TRACEABLE TO SRM186I AND 186II	YES		YES
	YES		
Sample ID			
Tested 11/04/21			
Tester RSWE			
Recorded 11/04/21			
PH NIST TRACEABLE TO SRM 187	YES		YES
	YES		
Sample ID			
Tested 11/04/21			
Tester RSWE			
Recorded 11/04/21			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

APPEARANCE AND COLOR	CLEAR YELLOW/ODORLESS APPEARANCE	PASS
Sample ID		
Tested	11/04/21	
Tester	RSWE	
Recorded	11/04/21	
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	T-IX,M81,M90,
Sample ID		
Tested	11/04/21	
Tester	RSWE	
Recorded	11/04/21	
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY	11/30/23
Sample ID		
Tested	11/04/21	
Tester	RSWE	
Recorded	11/04/21	
SHELF LIFE OF PRODUCT	2 YEARS TIME	2 YEARS
Sample ID		
Tested	11/04/21	
Tester	RSWE	
Recorded	11/04/21	
Lot #	1110173	
Made	11/04/21	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager