RICCA CHEMICAL COMPANY®

Certificate of Analysis

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25° C

	\mathbf{Lot}	Number:	1203959
--	----------------	---------	---------

Product Number: 1550

Manufacture Date: MAR 16, 2022 Expiration Date: MAR 2024

The certified value for this product is confirmed in independent testing by a second qualified chemist. The NIST traceable pH value is certified to ± 0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05 .

°C pH	$\begin{array}{c} 0 \\ 7.12 \end{array}$	$5 \\ 7.09$	$\begin{array}{c} 10 \\ 7.06 \end{array}$	$\begin{array}{c} 15 \\ 7.04 \end{array}$	$\begin{array}{c} 20 \\ 7.02 \end{array}$	$\begin{array}{c} 25 \\ 7.00 \end{array}$	$\begin{array}{c} 30 \\ 6.99 \end{array}$	$\begin{array}{c} 35 \\ 6.98 \end{array}$	$\begin{array}{c} 40 \\ 6.98 \end{array}$	$\begin{array}{c} 45 \\ 6.97 \end{array}$	$\begin{array}{c} 50 \\ 6.97 \end{array}$		
Name						CA	\S#			Grade			
Water						77	32-18-5			ACS/AS	STM/USP/I	ΞP	
Sodiun	n Hydro	oxide				13	10-73-2			Reagen	ıt		
Potass	ium Dil	nydroge	n Phosp	hate		77	78-77-0			ACS			
Sodiun	n Phosp	hate Di	basic			75	58-79-4			ACS			
Preser	vative					Pr	oprietar	'y					

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
pH at 25°C (Method: SQCP027, SQCP033)	6.990-7.010	7.004	186-I-g, 186-II-g, 191d

pH at 25°C (Method: SQCP027, SQCP033) 0.01 0.01 Uncertainty

pH measurements were performed in our Arlington, TX laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution.

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)			
1550-1	4 L natural poly	24 months			
1550-16	500 mL natural poly	24 months			
1550-32	1 L natural poly	24 months			
1550-5	20 L Cubitainer®	24 months			
Recommended Storage: 15°C - 30°C (59°F - 86°F)					

Jason Thomas (03/16/2022) Operations Manager This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."



This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.