

448 West Fork Dr Arlington, TX 76012 http://www.riccachemical.com 1-888-GO-RICCA customerservice@riccachemical.com

Certificate of Analysis

Buffer, Reference Standard, pH 8.00 ± 0.01 at 25° C

Lot Number: 1204954 Product Number: 1580

Manufacture Date: APR 14, 2022

Expiration Date: APR 2024

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST Traceable pH value is certified to ± 0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05 .

10 20 2530 35 40 15 45 8.00 8.00 7.98 8.03 8.00 7.99 7.98 7.98 pН 8.05

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	
Phosphoric Acid	7664-38-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
pH at 25°C (Method: SQCP027, SQCP033)	7.990-8.010	8.005	186-I-g, 186-II-g, 191d
pH at 25°C (Method: SQCP027, SQCP033) Uncertainty	0.02	0.02	

pH measurements were performed in our Arlington, TX laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution.

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1580-1	4 L natural poly	24 months
1580-16	500 mL natural poly	24 months
1580-2.5	10 L Cubitainer®	24 months
1580-32	1 L natural poly	24 months
1580-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Version: 1.3 **Lot Number:** 1204954 **Product Number:** 1580 Page 1 of 2

The

Jason Thomas (04/14/2022)

Operations Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials \cdots Contents of Certificates and Labels."



Certificate # L2387 Testing

 $This test \ report \ shall \ not \ be \ reproduced, \ except \ in \ full, \ without \ the \ written \ approval \ of \ Ricca \ Chemical \ Company.$

Version: 1.3 Lot Number: 1204954 Product Number: 1580 Page 2 of 2