

Certificate of Analysis

Potassium Permanganate, 0.100 Normal (N/10), 0.0200 Molar (M/50)

Lot Number: 1204C72

Product Number: 6400

Manufacture Date: MAY 03, 2022

Expiration Date: OCT 2023

The shorter shelf life of some sizes of this product is due to chemical incompatibility with the plastic packaging material. For these sizes, please consult the package label for expiration date. Product packaged in plastic containers or cubitainers should be restandardized before each use if extreme accuracy is required.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Permanganate	7722-64-7	

Test	Specification	Result	NIST SRM#
Appearance	Purple liquid	Passed	
Assay (vs. Sodium Oxalate)	0.0999-0.1001 N at 20°C	0.1001 N at 20°C	8040

Specification	Reference
0.1 N Potassium Permanganate Solution	APHA (3500-Mn B)
Potassium permanganate solution, standardized 0.1000 ± 0.0005 N	TAPPI (T 236 cm-85)
Potassium Permanganate, 0.1 N	EPA (218.3)
Potassium permanganate, 0.1 N	TAPPI (T 610 sp-97)
Potassium Permanganate, 0.1 N	EPA (245.2)
Potassium Permanganate, 0.1 N Volumetric Solution	ASTM (E 200)
Potassium Permanganate, 0.1 Normal Volumetric Solution	ACS (N/A)
0.02 M Potassium permanganate Volumetric Solution	EP (3005300)
Potassium Permanganate TS	USP (N/A)
Potassium Permanganate, Tenth-Normal (0.1 N) Volumetric Solution	USP (N/A)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6400-1	4 L amber glass	18 months
6400-16	500 mL amber glass	18 months
6400-32	1 L amber glass	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

A handwritten signature in black ink, appearing to read 'Jason Thomas', is displayed on a light gray rectangular background.

Jason Thomas (05/03/2022)

Operations Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.