



CERTIFICATE OF QUALITY

Purepoint 50uL Filter Pipette Tips, Sterile

September 5, 2019

Part Number	FPT-050
Lot Number	190706-361-2
Sterilization	
Product labeled as sterile is irradiated and dosimetrically released upon ISO 11137 recommended practices in effect at time of validation. Sterilization records are reviewed and signed off by qualified personnel for product release.	
Non-Pyrogenic	
Products labeled non-pyrogenic have been validated per FDA guidelines on LAL (Limulus Amebocyte Lysate) testing for medical devices and company guidelines. The acceptance level for product is less than 0.5 EU/mL.	
DNase & RNase Free	
This product has been tested and is free of any detectable Nucleic Acid.	
Trace Metal	
All material are certified metal-free by material manufacturers before they are turned into production which compliance with US Pharmacopoeia guidelines.	
ISO 9001:2008 Quality System Compliance	
Products are manufactured products are manufactured in accordance with ISO13485:2003 & ISO 9001:2008 quality management systems in a class 100,000 clean room.	
Quality Control Testing	
Products are manufactured in accordance with ISO 13485:2003 & ISO 9001:2008 quality management systems in a class 100,000 clean room.	

Filtrous Inc.
145 Commerce Way Suite A
Walnut, CA 91789
United States
info@filtrous.com
310-853-0388

Signed

Manager — Supplier Qualification
Filtrous Inc.