



Certificate of Analysis

Baird-Parker Agar
Catalog Number: B02-102
Lot Number: 191216022301
Expiration: December 2022

3651 Clipper Mill Rd.
Baltimore, MD 21211
Phone (410) 467-9983
Fax (410) 467-5088

Physical And Chemical Tests

<u>Test Parameter</u>	<u>Specification</u>	<u>Result</u>
DCM Appearance	Homogeneous, free flowing, beige	Homogeneous, free flowing, beige
PPM Appearance*	Opaque, canary yellow	Opaque, canary yellow
pH at 25 °C (6.0% solution)	7.0 +/- 0.2	6.96

Microbiological Tests

<u>Microorganisms</u>	<u>Growth*</u>	<u>Reactions*</u>
<i>Bacillus subtilis</i> ATCC 6633	Inhibition	N/A
<i>Enterococcus faecalis</i> ATCC 29212	Very Good Black col. w/o halo	Lipase & Lecithinase -
<i>Escherichia coli</i> ATCC 25922	Inhibited	N/A
<i>Proteus mirabilis</i> ATCC 12453	Fair Brown col. w/o halo	Lipase & Lecithinase -
<i>Staphylococcus aureus</i> ATCC 6538	Excellent Black col. w/ halo	Lipase & Lecithinase +
<i>Staphylococcus aureus</i> ATCC 25923	Excellent Black col. w/ halo	Lipase & Lecithinase +
<i>Staphylococcus epidermidis</i> ATCC 12228	Excellent Grey-Bk col. w/o halo	Lipase & Lecithinase -

* The following results were obtained from type cultures after incubation at a temperature of 37°C and observed after 24-48 hours. Media prepared with Egg Yolk Tellurite Enrichment.

Laboratory Results: Satisfactory

Country of Manufacture:
Testing performed by:
Completed on:
QC Release by:

USA
A.Kazmi/ H.McGinnis
6-15-20
Elizabeth Fox