



Certificate of Analysis

Buffered Peptone Water (ISO)
Catalog Number: B02-130
Lot Number: 201228012312
Expiration: December 2023

3651 Clipper Mill Rd.
Baltimore, MD 21211
Phone (410) 467-9983
Fax (410) 467-5088

Physical And Chemical Tests

DCM Appearance: (Homogeneous, free flowing, light beige)	Free flowing, homogeneous, light beige
PPM Color/Clarity: (Pale to light yellow/ Clear to brilliant/none to light ppt)	Light yellow/clear/none
pH: (2.0% solution at 25°C: 7.0 +/- 0.2):	7.03

Microbiological Tests

<u>Microorganisms</u>	<u>Growth*</u>
<i>Escherichia coli</i> ATCC 25922	Excellent
<i>Salmonella enteritidis</i> ATCC 13076	Excellent
<i>Salmonella typhi</i> ATCC 19430	Very Good
<i>Salmonella typhimurium</i> ATCC 14028	Excellent

* The following results were obtained from type cultures after incubation at a temperature of 35 °C and observed after 18-24 hours.

Laboratory Results: Satisfactory

Country of Manufacture:	USA
Testing performed by:	B. Henson/Q. McClary
Completed on:	01-06-21
QC Release by:	Abida Kazmi