



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 1460			
BUFFER REFERENCE pH 4.0/RED			
Test	Target/UOM	Range	Result
PH OF PRODUCT/SOLUTION	4.008	3.998 - 4.018	4.004
	PH		
Sample ID			
Tested 05/02/22			
Tester BPHI			
Recorded 05/02/22			
PH OF PRODUCT/SOLUTION	4.008	3.998 - 4.018	4.004
	PH		
Sample ID			
Tested 05/10/22			
Tester BPHI			
Recorded 05/10/22			
QCSP 20 PH CALIBRATION	YES		YES
METHODS	METHOD		
Sample ID			
Tested 05/02/22			
Tester BPHI			
Recorded 05/02/22			
QCSP 20 PH CALIBRATION	YES		YES
METHODS	METHOD		
Sample ID			
Tested 05/10/22			
Tester BPHI			
Recorded 05/10/22			
Actual Temp	25.0	24.0 - 26.0	25.0
	DEG C		
Sample ID			
Tested 05/02/22			
Tester BPHI			
Recorded 05/02/22			
Actual Temp	25.0	24.0 - 26.0	25.0
	DEG C		
Sample ID			
Tested 05/10/22			
Tester BPHI			
Recorded 05/10/22			
pH Traceable to N.I.S.T. SRM'S	YES		YES
	YES		
Sample ID			
Tested 05/02/22			
Tester BPHI			
Recorded 05/02/22			
pH Traceable to N.I.S.T. SRM'S	YES		YES
	YES		
Sample ID			
Tested 05/10/22			
Tester BPHI			
Recorded 05/10/22			



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

<p>PH NIST TRACEABLE TO SRM 185</p> <p>Sample ID Tested 05/02/22 Tester BPHI Recorded 05/02/22</p> <p>PH NIST TRACEABLE TO SRM 185</p> <p>Sample ID Tested 05/10/22 Tester BPHI Recorded 05/10/22</p> <p>PH NIST TRACEABLE TO SRM 187</p> <p>Sample ID Tested 05/02/22 Tester BPHI Recorded 05/02/22</p> <p>PH NIST TRACEABLE TO SRM 187</p> <p>Sample ID Tested 05/10/22 Tester BPHI Recorded 05/10/22</p> <p>NIST TRACEABLE TO SRM186I AND 186II</p> <p>Sample ID Tested 05/02/22 Tester BPHI Recorded 05/02/22</p> <p>NIST TRACEABLE TO SRM186I AND 186II</p> <p>Sample ID Tested 05/10/22 Tester BPHI Recorded 05/10/22</p> <p>APPEARANCE AND COLOR</p> <p>Sample ID Tested 05/02/22 Tester BPHI Recorded 05/02/22</p> <p>APPEARANCE AND COLOR</p> <p>Sample ID Tested 05/10/22 Tester BPHI Recorded 05/10/22</p>	<p>YES YES</p> <p>YES YES</p> <p>YES YES</p> <p>YES YES</p> <p>YES YES</p> <p>YES YES</p> <p>YES YES</p> <p>YES YES</p> <p>CLEAR RED/SLIGHT THYMOL SMELL APPEARANCE</p> <p>CLEAR RED/SLIGHT THYMOL SMELL APPEARANCE</p>	<p>YES</p> <p>YES</p> <p>YES</p> <p>YES</p> <p>YES</p> <p>YES</p> <p>YES</p> <p>YES</p> <p>CLEAR RED/SLIGHT THYMOL SMELL</p> <p>PASS</p>
---	---	--



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M-56
Sample ID		
Tested 05/02/22		
Tester BPHI		
Recorded 05/02/22		
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT	M81,(O)T-IX,M56
Sample ID		
Tested 05/10/22		
Tester BPHI		
Recorded 05/10/22		
EXPIRATION DATE	MM/DD/YY	04/30/24
	MM/DD/YYYY	
Sample ID		
Tested 05/02/22		
Tester BPHI		
Recorded 05/02/22		
EXPIRATION DATE	MM/DD/YY	04/30/24
	MM/DD/YYYY	
Sample ID		
Tested 05/10/22		
Tester BPHI		
Recorded 05/10/22		
SHELF LIFE	2 YEARS TIME	2 YEARS
Sample ID		
Tested 05/02/22		
Tester BPHI		
Recorded 05/02/22		
SHELF LIFE	2 YEARS TIME	2 YEARS
Sample ID		
Tested 05/10/22		
Tester BPHI		
Recorded 05/10/22		
Lot #	2040687	
Made	05/10/22	

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536
Website: www.aquasolutions.org

Page: 4

Date: 05/10/22 at 7:32 AM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager