



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer: DEP040 / SCI (SCIENTIFIC CONSUMABLES&INSTRU)

Packaged Product: 1525-10L BUFFER REFERENCE pH 10.0/BLUE (10LC)			
Lot # 2050993			
Chemical Product: 1525 BUFFER REFERENCE pH 10.0/BLUE			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
PH OF PRODUCT/SOLUTION	10.000	9.990 - 10.010	10.008
+ - PH Measurement Uncertainty	0.014	= 0.014	0.014
QCSP 20 PH CALIBRATION METHODS	YES		YES
Actual Temp	25.0	24.9 - 25.1	25.0
pH Traceable to N.I.S.T. SRM'S	YES		YES
PH NIST TRACEABLE TO SRM 185	YES		YES
NIST TRACEABLE TO SRM186I AND 186II	YES		YES
PH NIST TRACEABLE TO SRM 187	YES		YES
TRACEABLE TO NIST SRM 191I AND SRM 191II	YES		YES
APPEARANCE AND COLOR	CLEAR BLUE/ODORLESS		PASS
INSTRUMENTS USED DURING PREPARATION	APPEARANCE INSTRUMENT		M81,M56,T-IX,
EXPIRATION DATE	MM/DD/YY		05/30/24
SHELF LIFE OF PRODUCT	MM/DD/YYYY		
	2 YEARS		2 YEARS
	TIME		
Made 05/12/22			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.



Aqua Solutions, Inc.
Ph: 281-479-2569
Fax: 281-479-2790
P.O. Box 70
Deer Park, TX 77536

Website: www.aquasolutions.org

Page: 2

Date: 11/11/22 at 12:48 PM

**Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569**

Certificate of Analysis / QC Results

Customer: DEP040 / SCI (SCIENTIFIC CONSUMABLES&INSTRU)

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager