



Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

10/31/2020 18:44:26 GMT
Page 1 of 1

Customer Name: Conbar Processing Facility: Salem Work Order # 2767708
P.O.# n/a Sales Order # 2624618
10.0 - 40.0 kGy Conbar, Gamma Treatment Irradiation Date/Time: 10/30/2020 11:34:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2052510	5665S
102.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2052510	8816-50/5665S
103.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2052510	882-24/2405-25S
104.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2052510	2440-25S/2405-25S
105.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2052510	8832-25 scoop
106.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2052510	8832-25 scoop
107.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2052510	8832-25 scoop
108.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2052510	8832-25 scoop
	8	EA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	10.0 kGy	40.0 kGy	14.2 KGY	Pass	SBARBIERI	10/30/2020 15:44:40 GMT
		Reason Code Test				STEVEN BARBIERI	
450.00	Maximum Dose	10.0 kGy	40.0 kGy	25.1 KGY	Pass	SBARBIERI	10/30/2020 15:44:49 GMT
		Reason Code Test				STEVEN BARBIERI	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: HEATHER ZMIJEWSKI
Reason: Work Order Completions

Date: 10/31/2020 18:41:40 GMT