



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>Packaged Product:</b> 8810			
SODIUM THIOSULFATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
<b>PREPARED TO FORMULATION ON FILE</b>	YES		YES
Sample ID	YES		
Tested 06/17/22			
Tester SNEL			
Recorded 06/17/22			
<b>PH OF PRODUCT/SOLUTION</b>	7.000	6.000 - 12.500	10.229
	PH		
Sample ID			
Tested 06/17/22			
Tester SNEL			
Recorded 06/17/22			
<b>Actual Normality</b>	0.1000	0.0995 - 0.1005	0.1005
	EQ/L		
Sample ID			
Tested 06/17/22			
Tester SNEL			
Recorded 06/17/22			
<b>QCSP10:STANDARDIZATION OF SODIUM THIOSUL</b>	YES		YES
	QCSP10		
Sample ID			
Tested 06/17/22			
Tester SNEL			
Recorded 06/17/22			
<b>QCSP35:STANDARDIZATION OF SODIUM THIOSUL</b>	YES		YES
	QCSP35		
Sample ID			
Tested 06/17/22			
Tester SNEL			
Recorded 06/17/22			
<b>APPEARANCE AND COLOR</b>	CLEAR WATER WHITE		PASS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 06/17/22			
Tester SNEL			
Recorded 06/17/22			
<b>INSTRUMENTS USED DURING PREPARATION</b>	INSTRUMENT		M85,M100,T156,M55,M81,BID0 0787,TUS
Sample ID			
Tested 06/17/22			
Tester SNEL			
Recorded 06/17/22			
<b>DATE OF PREPARATION</b>	MM/DD/YYYY		06/17/22
Sample ID			
Tested 06/17/22			
Tester SNEL			
Recorded 06/17/22			



**Aqua Solutions, Inc.**  
**6913 Highway 225**  
**Deer Park, TX, 77536**  
**281/479-2569**  
**Certificate of Analysis / QC Results**  
**Customer PO # :**

<b>EXPIRATION DATE</b>	MM/DD/YY		12/30/22
	MM/DD/YYYY		
Sample ID			
Tested	06/17/22		
Tester	SNEL		
Recorded	06/17/22		
<b>SHELF LIFE</b>	6 MONTHS		6 MONTHS
	TIME		
Sample ID			
Tested	06/17/22		
Tester	SNEL		
Recorded	06/17/22		
<b>TRACEABLE TO NIST SRM 136</b>	YES		YES
	YES		
Sample ID			
Tested	06/17/22		
Tester	SNEL		
Recorded	06/17/22		
<b>Actual Density</b>	1.01127	1.01080 - 1.01173	1.01141
	GM/ML		
Sample ID			
Tested	06/17/22		
Tester	SNEL		
Recorded	06/17/22		
<b>Lot #</b>	2060803		
<b>Made</b>	06/17/22		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

  
**Quality Manager**