



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536

Website: www.aquasolutions.org

Page: 1

Date: 10/11/22 at 12:47 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer: DEP040 / SCI (SCIENTIFIC CONSUMABLES&INSTRU)

Packaged Product: 4750-4L IODINE 0.1 NORMAL SOLN. CERTIFIED (4L)			
Lot # 2080439			
Chemical Product: 4750 IODINE 0.1 NORMAL SOLN. CERTIFIED			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Actual Normality	0.1000	0.0995 - 0.1005	0.0997
APPEARANCE AND COLOR	EQ/L DARK BROWN LIQUID		PASS
INSTRUMENTS USED DURING PREPARATION	APPEARANCE INSTRUMENT		M102,M103,T-142,M55,M81,
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		08/30/23
SHELF LIFE OF PRODUCT	1 YEAR		1 YEAR
Actual Density	TIME 1.03678	1.03551 - 1.03805	1.03646
QCSP12:STANDARDIZATION OF IODINE SOLUTIO	GM/ML YES QCSP12		YES
Made 08/25/22			

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager