



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer: DEP040 / SCI (SCIENTIFIC CONSUMABLES&INSTRU)

Packaged Product: 0531-4LC
 AMMONIUM THIOCYANATE 0.1 NORMAL (4LC)

Lot # 2082307

Chemical Product: 0531
 AMMONIUM THIOCYANATE 0.1 NORMAL

Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Actual Normality	0.1000 EQ/L	0.0995 - 0.1005	0.0997
QCSP13:STANDARDIZATION	YES		YES
POTASSIUM&AMMONIU APPEARANCE AND COLOR	QCSP13 CLEAR LIQUID		PASS
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M83,(O)T-34,M55,BID3902,PID 363
EXPIRATION DATE	MM/DD/YY MM/DD/YYYY		09/30/23
SHELF LIFE OF PRODUCT	SHELF LIFE TIME		1 YEAR
Actual Density	1.00002 GM/ML	0.99994 - 1.00009	1.00004

Made 09/20/22

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager