



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536

Website: www.aquasolutions.org

Page: 1

Date: 08/31/22 at 2:02 PM

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569

Certificate of Analysis / QC Results

Customer: DEP040 / SCI (SCIENTIFIC CONSUMABLES&INSTRU)

Packaged Product:	THE294 CONDUCTIVITY STANDARD 575 UMHOS @25 C (1L)		
Lot #	2082462		
Chemical Product:	THE294F CONDUCTIVITY STANDARD 575 UMHOS @25 C		
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
APPEARANCE AND COLOR	CLEAR LIQUID APPEARANCE		PASS
QCSP30:CONDUCTIVITY STD&MEASUREMENT	YES QCSP30		YES
TRACEABLE TO N.I.S.T. SRM 3191	YES		YES
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M92,T234
CONDUCTIVITY (US/CM)	575 US/CM	573 - 577	575
Made	08/29/22		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.

Quality Manager