



Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

Packaged Product: 8810			
SODIUM THIOSULFATE 0.1 NORMAL			
Test	Target/UOM	Range	Result
PREPARED TO FORMULATION ON FILE	YES		YES
Sample ID	YES		
Tested 11/01/22			
Tester CCOA			
Recorded 11/01/22			
PH OF PRODUCT/SOLUTION	7.000	6.000 - 12.500	9.799
	PH		
Sample ID			
Tested 11/01/22			
Tester CCOA			
Recorded 11/01/22			
Actual Normality	0.1000	0.0995 - 0.1005	0.0999
	EQ/L		
Sample ID			
Tested 11/01/22			
Tester CCOA			
Recorded 11/01/22			
QCSP10:STANDARDIZATION OF SODIUM THIOSUL	YES		YES
	QCSP10		
Sample ID			
Tested 11/01/22			
Tester CCOA			
Recorded 11/01/22			
QCSP35:STANDARDIZATION OF SODIUM THIOSUL	YES		YES
	QCSP35		
Sample ID			
Tested 11/01/22			
Tester CCOA			
Recorded 11/01/22			
APPEARANCE AND COLOR	CLEAR WATER WHITE		PAS
	LIQUID		
	APPEARANCE		
Sample ID			
Tested 11/01/22			
Tester CCOA			
Recorded 11/01/22			
INSTRUMENTS USED DURING PREPARATION	INSTRUMENT		M87,M83,T-72,M79,M91,
Sample ID			
Tested 11/01/22			
Tester CCOA			
Recorded 11/01/22			
DATE OF PREPARATION	MM/DD/YYYY		11/01/22
Sample ID			
Tested 11/01/22			
Tester CCOA			
Recorded 11/01/22			



Aqua Solutions, Inc.
 Ph: 281-479-2569
 Fax: 281-479-2790
 P.O. Box 70
 Deer Park, TX 77536
 Website: www.aquasolutions.org

Aqua Solutions, Inc.
6913 Highway 225
Deer Park, TX, 77536
281/479-2569
Certificate of Analysis / QC Results
Customer PO # :

EXPIRATION DATE	MM/DD/YY		05/30/23
	MM/DD/YYYY		
Sample ID			
Tested	11/01/22		
Tester	CCOA		
Recorded	11/01/22		
SHELF LIFE	6 MONTHS		6 MONTHS
	TIME		
Sample ID			
Tested	11/01/22		
Tester	CCOA		
Recorded	11/01/22		
TRACEABLE TO NIST SRM 136	YES		YES
	YES		
Sample ID			
Tested	11/01/22		
Tester	CCOA		
Recorded	11/01/22		
Actual Density	1.01127	1.01080 - 1.01173	1.01134
	GM/ML		
Sample ID			
Tested	11/01/22		
Tester	CCOA		
Recorded	11/01/22		
Lot #	2100402		
Made	11/01/22		

This is to certify that the product listed above has been prepared according to the agreed upon formulation. The solutions producer has a Quality Management System which governs each step of the manufacturing process to insure the production of a consistent product. Traceability from the producer's lot numbers to the original manufacturer's lot number is maintained. Certificates of Analysis for the raw materials used to prepare this product are available upon request.

The weights and/or volumes used to prepare this product are N.I.S.T. Traceable. All balances used in the preparation of product are calibrated daily against N.I.S.T. Traceable weights. The balances are maintained and serviced on a regular basis by an outside certified company. All volumetric glassware used is N.I.S.T. Traceable and certified as meeting Class A specifications.

Unless otherwise agreed to, all chemicals used in the preparation of this product are Reagent ACS grade.


Quality Manager