



625 East Bunker Court
Vernon Hills, IL 60061
Ph: 800-886-8675 / 847-622-0456
argos-tech.com

Product Name: Pipettes
Description: Polystyrene (PS) Serological
Pipettes, 25 mL, Sterile Indv; 150/Cs
Manufacturer Part Number: 13002-19

Lot Number: 210107-248
Date of Manufacture: 01/07/21
Manufacture Expiration Date: 01/07/2024
Country of Origin: China

Quality System Compliance: Products are manufactured under the ISO 9001:2008 & ISO13485:2003 standard. Products are Inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

Sterilization Process: Product labeled as sterile is irradiated. The specified dose of irradiation for this item is within the range of 12.0 kGy to 24.0 kGy.

Except as otherwise set forth in writing by Argos, there are no other warranties, express or implied, including without limitation warranties of marketability or fitness for a particular purpose, use or application, that apply to this certificate or the subject product. It is the responsibility of the user to determine the suitability of the subject product for any particular use or application.

Argos Certificates Team

This information is subject to change and be considered the current version only on the day of print.

Date Printed: 04/29/2021

Page 1 of 1