



Certificate of Analysis

Sabouraud Dextrose Agar
Catalog Number: S19-100
Lot Number: 210108022501
Expiration: January 2025

3651 Clipper Mill Rd.
Baltimore, MD 21211
Phone (410) 467-9983
Fax (410) 467-5088

Physical And Chemical Tests

<u>Test Parameter</u>	<u>Specification</u>	<u>Result</u>
DCM Appearance	Homogeneous, free flowing, beige	Homogeneous, free flowing, beige
PPM Appearance	Trace to slight haze, yellow to light amber	Slight haze, light amber
pH at 25 °C (6.5% solution)	5.6 +/- 0.2	5.75

Microbiological Tests

<u>Microorganisms</u>	<u>Expected Growth & Appearance</u>	<u>Results*</u>
<i>Aspergillus brasiliensis</i> (formerly <i>A.niger</i>) ATCC 16404	Good to excellent White cottony, black spores	Excellent Wh. cottony, Bk spores
<i>Candida albicans</i> ATCC 10231	Good to excellent Off white, pasty texture	Very Good Off wh. pasty
<i>Saccharomyces cerevisiae</i> ATCC 9763	Good to excellent Off white, pasty texture	Very Good Off wh. pasty
<i>Trichophyton mentagrophytes</i> ATCC 9533	Good to excellent White cottony	Excellent Wh. cottony

* The following results were obtained from type cultures after incubation at a temperature of 25 -30°C and observed after 2-7 days.

Laboratory Results: Satisfactory

Country of Manufacture	USA
Testing performed by:	Q. McClary/B. Henson
Completed on:	1-27-21
QC Release by:	Abida Kazmi