1841 Broad Street Pocomoke City, MD 21851 http://www.riccachemical.com 1-888-GO-RICCA customerservice@riccachemical.com

Certificate of Analysis

Buffer, Reference Standard, pH 4.00 ± 0.01 at 25°C (Color Coded Red)

Lot Number: 2104H21

Product Number: 1501

Manufacture Date: APR 29, 2021

Expiration Date: APR 2023

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST Traceable pH value is certified to ± 0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05 .

10 15 20 25 30 40 45 Hq 50 4.00 4.00 4.00 4.00 4.00 4.00 4.01 4.02 4.03 4.04 4.06

	1.00
CAS#	Grade
Proprietary	Purified
7732-18-5	ACS/ASTM/USP/EP
877-24-7	Buffer
Proprietary	Commercial
	7732-18-5 877-24-7

Test	Specification	Result	NIST SRM#
Appearance	Red liquid	Passed	- IDI DIWIF
pH at 25°C (Method: SQCP027, SQCP033)	3.990-4.010	4.000	185i, 186-I-g,
pH at 25°C (Method: SQCP027, SQCP033) Uncertainty	0.01	0.01	186-II-g

Specification		
	Reference	12/16
Commercial Buffer Solutions Buffer B	ASTM (D 1293 B)	
Buffer B	ASTM (D 5464)	
pH measurements were perferred in a	ASTM (D 5128)	

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

	101 oddi fot manufactured.	
Size / Package Type	Shelf Life (Unopened Container)	
20 x 20 mL pack		
	24 months	
	Size / Package Type 20 x 20 mL pack 20 mL pouch (59°F - 86°F)	

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Version: 1.3

Lot Number: 2104H21 Product Number: 1501 Page 1 of 2

Kerry Kengbray

Kerry Kingsbury (04/29/2021)

Quality Control Supervisor

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."



This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3

Lot Number: 2104H21

Product Number: 1501