

Quality Certificate



229224B

10mL Open End Serological Pipets Individually Wrapped Pipets, Sterile



Lot Number: 211017-624-A
Expiration Date: 10/17/2024

This product is assembled in accordance with ISO 9001:2015 and ISO 13485:2016 quality management systems in a class 100,000 clean room.

Component Materials

Virgin Polystyrene, meets USP, Class VI requirements.

Sterilization Assurance Level

SAL 10^{-6}

This lot has been gamma irradiated and dosimetrically released. Procedures for determining the dosage comply with ANSI/AAMI/ISO 11137.

Cytotoxicity

All material resins are qualified using USP and/or ISO 10993 standards for cytotoxicity and have been shown to be non toxic.

Non-pyrogenic

Products have been validated to have less than 0.1 EU/mL.

RNase/DNase Testing

This product has been tested and is free of any detectable RNase/DNase contamination.

Quality Control Testing

Representative Samples have been selected and inspected according to ANSI/ASQ Z1.4 Sampling Procedures and Tables for Inspection for Attributes. This lot has passed inspection.

Quality Control Testing

CELLTREAT serological pipets are accurate to +/- 2% at full volume.

Country of Origin

China

This manufacturing lot has met CELLTREAT quality standards and has been accepted by quality assurance.

Product Quality Department
CELLTREAT Scientific Products, LLC

CELLTREAT is a registered trademark of CELLTREAT Scientific Products, LLC.

20 Mill Street, Unit 130, Pepperell, MA 01463
Phone: 877-345-4620 | **Email:** service@celltreat.com