

## **CERTIFICATE OF ANALYSIS**

Product Name Methanol

Grade ACS/USP/NF/HPLC-UV Grade

Catalog # 33900HPLC Item # 101183

Batch # 211213-B009655

Date of Manufacture: 12/09/2021

Recommended Retest Date: 12/08/2026

Customer PO # P23610

Packaging Type Case Poly Fluor 4x1 Gal

| TEST                                    | MONO-<br>GRAPH | SPECIFICATION     | RESULT           | UNITS |
|---|----------------|-------------------|------------------|-------|
| Carbonyl Compounds -                    | ACS            | 0.001% max        | LT 0.001%        | N/A   |
| Acetaldehyde                            |                |                   |                  |       |
| Carbonyl Compounds - Acetone            | ACS            | 0.001% max        | None<br>Detected | N/A   |
| Appearance                              | ACS            | To Pass Test      | Pass             | N/A   |
| Assay (corrected for water)             | ACS            | 99.8% min         | 99.94            | %     |
| Color (APHA)                            | ACS            | 10 max            | 1                | N/A   |
| Carbonyl Compounds -                    | ACS            | 0.001% max        | LT 0.001%        | N/A   |
| Formaldehyde                            |                |                   |                  |       |
| Residue on Evaporation                  | ACS            | 0.001% max        | 0.000            | %     |
| Solubility in Water                     | ACS            | To Pass Test      | Pass             | N/A   |
| Substances Darkened by<br>Sulfuric Acid | ACS            | To Pass Test      | Pass             | N/A   |
| Substances Reducing                     | ACS            | To Pass Test      | Pass             | N/A   |
| Permanganate                            |                |                   |                  |       |
| Titrable Acid                           | ACS            | 0.0003 meq/g max. | 0.0002           | meq/g |
| Titrable Base                           | ACS            | 0.0002 meq/g max. | 0.0001           | meq/g |
| Water                                   | ACS            | NMT 0.1%          | 0.05             | %     |
| UV Absorption @205nm                    | ACS            | 1.00 max.         | 0.88             | N/A   |
| UV Absorption @210nm                    | ACS            | 0.80 max.         | 0.49             | N/A   |

www.pharmco.com | www.greenfield.com

Form: CofA-Standard, Rev 1.6, 04/13/22, RAD



## **CERTIFICATE OF ANALYSIS**

| TEST                                      | MONO-<br>GRAPH | SPECIFICATION                                | RESULT    | UNITS |
|---|----------------|--|-----------|-------|
| UV Absorption @220nm                      | ACS            | 0.40 max.                                    | 0.23      | N/A   |
| UV Absorption @230nm                      | ACS            | 0.20 max.                                    | 0.11      | N/A   |
| UV Absorption @240nm                      | ACS            | 0.10 max.                                    | 0.05      | N/A   |
| UV Absorption @260nm                      | ACS            | 0.04 max.                                    | 0.01      | N/A   |
| UV Absorption @280nm-400<br>nm            | ACS            | 0.01 max.                                    | 0.00      | N/A   |
| Acetone and Aldehydes (as Acetone)        | NF             | NMT 0.003%                                   | LT 0.003% | N/A   |
| Acidity                                   | NF             | NMT 0.45mL 0.020N NaOH required              | 0.20      | ml    |
| Alkalinity (as NH4)                       | NF             | NMT 0.20mL 0.020N H2SO4 required (3 ppm max) | 0.10      | ml    |
| Assay                                     | NF             | NLT 99.5%                                    | 100.00    | %     |
| Identification A (Infrared<br>Absorption) | NF             | To Pass Test                                 | Pass      | N/A   |
| Identification B (GC Analysis)            | NF             | To Pass Test                                 | Pass      | N/A   |
| Non -Volatile Residue                     | NF             | NMT 2mg (0.001% w/w)                         | 0         | mg    |
| Readily Carbonizable<br>Substances        | NF             | To Pass Test                                 | Pass      | N/A   |
| Readily Oxidizable Substances             | NF             | To Pass Test                                 | Pass      | N/A   |
| Water                                     | NF             | NMT 0.1%                                     | 0.05      | %     |

## **Certification and Compliance Statements**

This lot of Methanol complies with all of the current requirements listed in the United States Pharmacopeia, National Formulary and American Chemical Society monographs

This product is manufactured for Routine HPLC Analysis and meets the requirements for General Use HPLC Grade, ACS Monographs. This product is not intended for GC or critical HPLC analysis. See Distilled Grade for those applications.

www.pharmco.com | www.greenfield.com

Form: CofA-Standard, Rev 1.6, 04/13/22, RAD



## **CERTIFICATE OF ANALYSIS**

No chemicals whatsoever are used as solvents at any point in the manufacture, processing or packaging of Methanol. Only Class 2 and Class 3 residual solvents may appear as impurities / related substances / low level contaminants in Methanol. Concentration of Class 2 Option 1 and Class 3 residual solvents is below limits in the current USP/NF General Chapter <467>.

This product is not derived, nor does it come in contact with, any materials derived from bovine or other animal sources.

This product is for further commercial manufacturing, laboratory or research use, and may be used as a process solvent for pharmaceutical purposes. It is not intended for use as an active ingredient in drug manufacturing nor as a medical device or disinfectant. Appropriate/legal use of this product is the responsibility of the user.

This document was electronically signed by Gabriella Scoca on 12/15/2021 3:09:13 PM to indicate Quality Assurance Approval and to release this batch.

Form: CofA-Standard, Rev 1.6, 04/13/22, RAD