



Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

06/28/2021 14:02:15 GMT
Page 1 of 1

Customer Name:	Conbar	Processing Facility:	Salem	Work Order #	2912894
P.O.#	N/A			Sales Order #	2802642
	10.0 - 40.0 kGy	Conbar, Gamma Treatment	Irradiation Date/Time:		06/26/2021 01:24:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152506	2405-25S
102.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152506	8864-24/2400-25S
103.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152506	8816-50/5665S
104.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152506	8882-24/5665S
105.000	1	EA	Scoops 343-1099	343-1099-BL-40-48-75	2152506	5665S
106.000	1	EA	Scoops 343-1099	343-1099-BL-40-48-75	2152506	2440-25S/2410-25S
107.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152506	8832-25 scoop
108.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152506	8832-25 scoop
109.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152506	8832-25 scoop
110.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152506	8832-25 scoop
111.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152506	8882-24/2405-25S
	11	EA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	10.0 kGy	40.0 kGy	12.4 KGY	Pass	JNEWMAN	06/26/2021 09:16:23 GMT
450.00	Maximum Dose	Reason Code Test 10.0 kGy	40.0 kGy	28.8 KGY	Pass	Jeffrey Newman JNEWMAN	06/26/2021 09:16:35 GMT
		Reason Code Test				Jeffrey Newman	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: AMY L. EDWARDS
Reason: Work Order Completions

Date: 06/28/2021 14:01:46 GMT