



Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
 TEL 856 935-1430 FAX 856 935-0542 www.sterigenics.com

R55480102

08/27/2021 12:38:05 GMT
 Page 1 of 1

Customer Name: Conbar	Processing Facility: Salem	Work Order # 2949647
P.O.# N/A		Sales Order # 2847086
10.0 - 40.0 kGy	Conbar, Gamma Treatment	Irradiation Date/Time: 08/26/2021 18:48:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152508	8864-24/2410-25S
102.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152508	8816-50/2410-25S
103.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152508	2405-25S
104.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152508	8882-24
105.000	1	EA	Scoops 343-1099	343-1099-BL-40-48-75	2152508	2440-25S/5665S
106.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152508	8832-25 scoop
107.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152508	8832-25 scoop
108.000	1	EA	Scoops 148-637	148-637-BL-47-35-55	2152508	8832-25 scoop
	8	EA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	10.0 kGy	40.0 kGy	12.5 KGY	Pass	JNEWMAN Jeffrey Newman	08/27/2021 09:44:24 GMT
450.00	Maximum Dose	10.0 kGy	40.0 kGy	27.4 KGY	Pass	JNEWMAN Jeffrey Newman	08/27/2021 09:44:38 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: **AMY L. EDWARDS**
 Reason: **Work Order Completions**

Date: **08/27/2021 12:37:51 GMT**