

Certificate of Analysis

Hydrochloric Acid, 6.00 Normal

Lot Number: 2201910

Product Number: 3750

Manufacture Date: JAN 05, 2022

Expiration Date: DEC 2023

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless to slightly greenish-yellow liquid	Passed	
Assay (vs. Sodium Hydroxide/Phenolphthalein)	5.995-6.005 N at 20°C	6.005 N at 20°C	84

Specification	Reference
Hydrochloric Acid, 6 N	APHA (4500-S2- F)
Hydrochloric Acid, 6 N	APHA (4500-S2- E)
Hydrochloric Acid, 6 N	APHA (2530 C)
Hydrochloric Acid, 6 N	APHA (4500-NO3- E)
Hydrochloric acid, 6 M	TAPPI (T 244 om-93)
Hydrochloric acid, 6M	TAPPI (T 266 om-94)
Hydrochloric Acid, 6 N	EPA (376.1)
Hydrochloric Acid (6 N)	EPA (SW-846) (9030B)
Hydrochloric acid, 6 M	TAPPI (T 247 cm-83)
Dilute Hydrochloric Acid, 6 N	EPA (353.3)
Dilute Hydrochloric Acid, 6 N	EPA (353.2)
6 M Hydrochloric acid Volumetric Solution	EP (3001500)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
3750-1	4 L natural poly	24 months
3750-16	500 mL natural poly	24 months
3750-32	1 L natural poly	24 months
3750-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Myrlande Gilles (01/05/2022)

Quality Control Supervisor

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.