

Certificate of Analysis

Buffer, Reference Standard, pH 4.00 ± 0.01 at 25°C (Color Coded Red)

Lot Number:	2205B06
-------------	---------

Product Number: 1501

Manufacture Date: MAY 10, 2022 Expiration Date: APR 2024

The certified value for this product is confirmed in independent testing by a second qualified chemist. The NIST Traceable pH value is certified to ± 0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05 .

°C pH	$\begin{array}{c} 0 \\ 4.00 \end{array}$	$5\\4.00$	$\begin{array}{c} 10 \\ 4.00 \end{array}$	$\begin{array}{c} 15 \\ 4.00 \end{array}$	$\begin{array}{c} 20 \\ 4.00 \end{array}$	$\begin{array}{c} 25 \\ 4.00 \end{array}$	$\begin{array}{c} 30\\ 4.01 \end{array}$	$\begin{array}{c} 35\\ 4.02 \end{array}$	$\begin{array}{c} 40\\ 4.03\end{array}$	$\begin{array}{c} 45\\ 4.04\end{array}$	$50 \\ 4.06$	
Name						CA	\S#			Grade		
Water				7732-18-5			ACS/ASTM/USP/EP					
Potass	ium Ac	id Phtha	alate			87	7-24-7			Buffer		
Preservative				Proprietary			Commercial			 		
Red Dye				Proprietary			Purified			 		
Test				Specification			Result			NIST SRM#		
Appearance				Red liquid			Passed					
pH at 25°C (Method: SQCP027, SQCP033)			3.990-4.010			4.004			 185i, 186-I-g, 186-II-g			
-	pH at 25°C (Method: SQCP027, SQCP033) Uncertainty				0.02			0.02			 	

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer B	ASTM (D 5464)
Buffer B	ASTM (D 5128)

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty ir the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution.

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1501-20	20 x 20 mL pack	24 months
1501-20B	200 x 20 mL pack	24 months
R1501000-20F1	20 mL pouch	24 months
D 110, 1700		

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Muurkeef

Myrlande Gilles (05/10/2022) Quality Control This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."



Certificate # L2387.01 Testing

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.