

Quality Certificate

PTFE Syringe Filter,Pore Size:0.22µm, Diameter:25mm,Non-Sterile	Part Number: SFPTFE025022NBCU
	Lot Number: 297057253

ISO 9001 Quality Standard



The product was manufactured under Membrane Solutions Quality Management System and is approved by an accredited registering body to the ISO 9001 Quality Systems Standard.

Syringe Filter Lot Characterization

Membrane material	PTFE	Filtration area(cm ²)	2.98
Pore size	0.22µm	Holdup volume (µl)	<100
Wettability	Hydrophobic	Housing Material	Polypropylene
Inlet/Outlet	Female Luer Lock inlet and Male Luer slip outlet	Bubble point(psi)	23.2-26.1
Burst pressure(psi)	87.0	Flow rate(ml/min@10psi)	85-97

Bubble point and Flow rate were tested by ethyl alcohol.

HPLC Extractable Test (254nm)



We hereby certify that the above stated filters are in compliance with all applicable instructions. This certificate shall not be copied or reproduced without the written approval of Membrane Solutions corp.

Tested By: *[Signature]* 10-23-2021

QC Technician