

GENERAL LABORATORY PRODUCTS

Certificate of Analysis

8WM-1800-90
Letheen Broth
90 ml
36429
16-Jun-2021
Light to medium amber.
7.0 ± 0.2 at 25° C
2° - 30°C

Organism	Source	Results
Bacillus subtilis	ATCC 6633	Growth
Escherichia coli	ATCC 8739	Growth
Pseudomonas aeruginosa	ATCC 9027	Growth
Staphylococcus aureus	ATCC 6538	Growth

Sterility Check:

	Expected Results	Actual Results
Sterility	No Growth after 48 hrs @ 30-35°C	No growth
	Followed by No Growth after 72 hrs @ 20 - 25°C	No growth

General Laboratory Products, Inc.

4 Bluea

General Laboratory Products certifies that samples of this lot have met sterility, pH and performance standards. The pH may fluctuate within the stated range based upon the type of pH meter and probe used. The age of the product could have an effect on the variance of the pH. Growth promotion testing was performed according to USP guidelines, where applicable.