

Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

05/27/2021 20:46:30 GMT
Page 1 of 1

Customer Name: Bel-Art Products-NJ
P.O.# MD63173

Processing Facility: Salem

Work Order # 2891064
Sales Order # 2776065

23.5-50.0 kGy

BelArt, Gamma Treatment

Irradiation Date/Time:

05/27/2021 14:12:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	70	CA	369460000	STERILEWARE,SAMPLING SPOON,3 T	J350373	NA
102.000	10	CA	369220000(30)	STERILEWARE,SCOOP SAMPLING SYS	J350983	NA
103.000	20	CA	367400310	STRLWR,ECON SMPL SPN,PS 5ML	J351917	NA
104.000	8	CA	367400110	STRLWR,ECON SMPL SPN,PS 1.25ML	J352256	NA
105.000	60	CA	369440000(30)	STERILEWARE,SAMPLING SPOON,1 T	J350803	NA
106.000	70	CA	369460010	Tablespoon 1 0	J352180	NA
107.000	40	CA	369022002(20)	STERILWARE,SCOOP,RED,PS,2 OZ	J351598	NA
108.000	115	CA	369022010	STERILEWARE,SCOOP,RED,PS,2OZ,1	J351463	NA
	393	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	23.5 kGy	50.0 kGy	26.9 KGY	Pass	SBARBIERI	05/27/2021 19:09:21 GMT
450.00	Maximum Dose	Reason Code Test 23.5 kGy	50.0 kGy	39.8 KGY	Pass	STEVEN BARBIERI SBARBIERI	05/27/2021 19:09:31 GMT
		Reason Code Test				STEVEN BARBIERI	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: AMY L. EDWARDS
Reason: Work Order Completions

Date: 05/27/2021 20:46:11 GMT