



Certificate of Analysis

Product Number: 8WM-1800-90
Product Description: Lethen Broth
Volume (ML): 90 ml
Lot Number: 37114
Expiration Date: 11-Sep-2021
Color / Appearance Light to medium amber.
Final pH Range: 7.0 ± 0.2 at 25° C
Storage Temperature: 2° - 30°C

Organism	Source	Results
Bacillus subtilis	ATCC 6633	Growth
Escherichia coli	ATCC 8739	Growth
Pseudomonas aeruginosa	ATCC 9027	Growth
Staphylococcus aureus	ATCC 6538	Growth

Sterility Check:

Sterility	Expected Results	Actual Results
	No Growth after 48 hrs @ 30-35°C Followed by No Growth after 72 hrs @ 20 - 25°C	No growth No growth

General Laboratory Products, Inc.

General Laboratory Products certifies that samples of this lot have met sterility, pH and performance standards. The pH may fluctuate within the stated range based upon the type of pH meter and probe used. The age of the product could have an effect on the variance of the pH. Growth promotion testing was performed according to USP guidelines, where applicable.