## RICCA CHEMICAL COMPANY®

## **Certificate of Analysis**

## Buffer, Reference Standard, pH $10.00 \pm 0.01$ at 25°C (Color Coded Blue)

Lot	Number:	4202048
-----	---------	---------

**Product Number**: 1601

Manufacture Date: FEB 02, 2022 Expiration Date: JUL 2023

The certified value for this product is confirmed in independent testing by a second qualified chemist. The NIST traceable pH value is certified to  $\pm 0.01$  at 25 °C only. All other pH values at their corresponding temperatures are accurate to  $\pm 0.05$ .

	reaction of the				• •••••			p			
°C pH	0 10.31	$5 \\ 10.23$	$\begin{array}{c} 10\\ 10.17\end{array}$	$\begin{array}{c} 15\\ 10.11 \end{array}$	$\begin{array}{c} 20\\ 10.05 \end{array}$	$\begin{array}{c} 25\\ 10.00 \end{array}$	$\begin{array}{c} 30\\ 9.95\end{array}$	$35 \\ 9.91$	$\begin{array}{c} 40\\ 9.87\end{array}$	50 9.81	
Name						CA	S#			Grade	
Water						773	32-18-5			ACS/ASTM/USP/EP	
Sodium	Carbona	ite				497	7-19-8			ACS	
Sodium	Bicarbo	nate				144	1-55-8			ACS	
Sodium	Hydroxi	de				131	0-73-2			Reagent	
Preserv	ative					Pro	prietary				
Blue Dy	ye					Pro	prietary				

Test	Specification	Result	NIST SRM#
Appearance	Blue liquid	Passed	
pH at 25°C (Method: SQCP027, SQCP033)	9.990-10.010	10.005	186-I-g, 186-II-g, 191d
pH at 25°C (Method: SQCP027, SQCP033)	0.02	0.02	

Uncertainty

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer C	ASTM (D 5464)
Buffer C	ASTM (D 5128)

L2387.02) Batesville, IN under ISO/IEC 17025 accreditation (ANAB Certificate pН measurements were performed in our laboratory and Institute Standards and Technology (NIST) Standard Reference Material indicated certified traceable to National of as above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution

Volumetric glassware complies with of ASTM E 288 and NIST Circular 434; Class А tolerance requirements it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before firstuse and recalibrated regularly a thermometer traceable to NIST standards. All products are with prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1601-1	4 L natural poly	18 months
1601-16	500 mL natural poly	18 months
1601-1CT	4 L Cubitainer®	18 months
1601-2.5	10 L Cubitainer®	18 months

1601-32	1 L natural poly	18 months
1601-5	20 L Cubitainer®	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Scon Riggle

Scott Riggle (02/02/2022) Plant Manager This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."



Certificate # L2387.02 Testing

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.