

# Certificate of Analysis

(Petri Dish)

Reports No: 200904

Types:  sampling inspection

final inspection

Name of Goods	Disposable Petri dish K1004-J	Specification	60*15mm
Lot number	1007/920102	Package	500pcs /ctn
Quantity	325000	Source	Sterilized Cabinet
Date of the sample	2020.09.10	Date of report	2020.09.17
Inspection basis	Q/HGS 005—2011 (Disposable Petri dish Corporation Standard)		

Inspection project	The standard stipulating	Check competence	Sample	Qualified quality level (AQL)	Inspection result	Conclusion	
Appearance quality	1.Petri dish should colorless and transparent, surface bright and clean, can't have obvious distorts, scrape, perforation, impurity, greasy dirt, air bubble, burr, etc. 2.The border should level smooth and have no burr and flash.	Special test level S-2	13	10		Re	PASS
				Ac	Re	0	
				3	4		
Physics performance	1.The wall should evenly thick; the lid and the body should cooperate well. 2.Should not distort when using under the environment which the Temperature is no more than 80℃.		20	2.5		Re	PASS
				Ac	Re	0	
				1	2		
Measure	1.Height of cap 2.Diameter of cap 3. Height of base 4. Diameter of base	13	10		Re	PASS	
			Ac	Re	0		
			3	4			
Sterilized inspection	1.According to the GB/ T 14233.1 standard, The remaining of EO shouldn't more than 10mg/kg after sterilized by EO Ten days later. 2.This test should be according with all the requirement of GB/T 14233.2.	20	2.5		Re	PASS	
			Ac	Re	0		
			1	2			
			All qualified		PASS		

Conclusion: This product is examined according to Q/HGS 005—2011 Disposable Petri Dish Corporation Standard. Inspection result accord with Stipulate. This batch of products are PASSED.

Sheet No. JY001-05-03

Inspected by: Fang xinyi

Checked by: Lv youyou

Zhejiang Gongdong Medical Technology co., Ltd.



# Certificate of Analysis

(Petri Dish)

Reports No: 200905

Types:  sampling inspection

final inspection

Name of Goods	Disposable Petri dish K1008	Specification	30*15mm
Lot number	1007/920102	Package	2000pcs /ctn
Quantity	200000	Source	Sterilized Cabinet
Date of the sample	2020.09.10	Date of report	2020.09.17
Inspection basis	Q/HGS 005—2011 (Disposable Petri dish Corporation Standard)		

Inspection project	The standard stipulating	Check competence	Sample	Qualified quality level (AQL)	Inspection result	Conclusion
Appearance quality	1.Petri dish should colorless and transparent, surface bright and clean, can't have obvious distorts, scrape, perforation, impurity, greasy dirt, air bubble, burr, etc. 2.The border should level smooth and have no burr and flash.	Special test level S-2	13	10	Re	PASS
				Ac	Re	
				3	4	
Physics performance	1.The wall should evenly thick; the lid and the body should cooperate well. 2.Should not distort when using under the environment which the Temperature is no more than 80℃.	Special test level S-2	20	2.5	Re	PASS
				Ac	Re	
				1	2	
Measure	1.Height of cap 2.Diameter of cap 3. Height of base 4. Diameter of base	Special test level S-2	13	10	Re	PASS
				Ac	Re	
				3	4	
Sterilized inspection	1.According to the GB/ T 14233.1 standard, The remaining of EO shouldn't more than 10mg/kg after sterilized by EO Ten days later. 2.This test should be according with all the requirement of GB/T 14233.2.	Special test level S-2	20	2.5	Re	PASS
				Ac	Re	
				1	2	
All qualified				PASS		

Conclusion: This product is examined according to Q/HGS 005—2011 Disposable Petri Dish Corporation Standard. Inspection result accord with Stipulate. This batch of products are PASSED.

Sheet No. JY001-05-03

Inspected by: Fang xinyi

Checked by: Lv youyou

Zhejiang Gongdong Medical Technology co., Ltd.

