



An ISO 9001 Certified Company

*Certificate of Analysis*

COMMODITY: **COD2 Mercury-Free COD Reagent 3-150 mg/l Range**

COMMODITY NUMBER: **2565025**

MANUFACTURE DATE:

DATE OF ANALYSIS:

LOT NUMBER: **A0061**

**3/5/2020**

**3/6/2020**

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Function Test: Determination of the COD in a 120 ppm COD standard prepared from KHP vs a reagent blank.	114 to 126 ppm	118.0 ppm
Leaking vials per sample of 50.	0 to 2 units	0.0 units
Vial to Vial Variability: Statistical sampling of reagent blank absorbance vs DI water, measured at 420 nm and controlled to a 99% confidence level.	0 to 15 ppm	6.5 ppm

The expiration date is Feb 2024

Certified by \_\_\_\_\_

Scott Als  
Analytical Services Chemist