



An ISO 9001 Certified Company

Certificate of Analysis

COMMODITY: **Sulfide 1 Reagent**
COMMODITY NUMBER: **181649**
LOT NUMBER: **A2087**

MANUFACTURE DATE:
4/2/2022

DATE OF ANALYSIS:
4/7/2022

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULTS</i>
Function Test: Absorbance of a standard with a sulfide concentration of 0.5 ppm measured in 1 inch cells vs the reagent blank.	0.899 to 1.121 abs	1.0660 abs
Function Test: Absorbance of a standard with a sulfide concentration of 1.0 ppm measured in 1 inch cells vs the reagent blank.	1.828 to 2.214 abs	1.9830 abs
Reagent Blank: Absorbance measured at 655 nm in 1 inch cells vs deionized water.	0 to 0.031 abs	0.0100 abs

The expiration date is Nov 2025

Certified by _____

Scott Als
Analytical Services Chemist