

HACH COMPANY



An ISO 9001 Certified Company

P.O.Box 389
Loveland, CO 80539
(970) 669-3050

Certificate of Analysis

Page 1

COMMODITY: **Sulfide 2 Reagent**
COMMODITY NUMBER: **181749**
LOT NUMBER: **A3059**

MANUFACTURE DATE:
3/3/2023

DATE OF ANALYSIS:
3/4/2023

<i>TEST</i>	<i>SPECIFICATIONS</i>	<i>RESULT:</i>
Potassium Dichromate Concentration	0.0288 to 0.0352 N	0.03130 N

The expiration date is Mar 2028

A handwritten signature in cursive script that reads "Scott Als".

Certified by _____

Scott Als
Analytical Services Chemist