



Certificate of Analysis

Description: SILVER AA STANDARD, 1000ppm (1mL = 1mg Ag)

Mfg. Date: 06/03/2019

Catalog Number: LC22420

Exp. Date: 06/03/2021

Lot Number: J149-24

Analytical Section

Test	Specification	Test Result
Appearance	clear, colorless solution	Pass Test
Concentration ppm Ag	1000ppm Ag +/- 10ppm	1000 ppm Ag
Concentration mg Ag/mL	1.000 +/- 0.010 mg Ag/mL	1.000 mg Ag/mL
Traceable to NIST	Potassium Chloride	999b

Intended Use - Product is intended for use in manufacturing procedures and laboratory procedures and protocols.

Storage Information - Unless otherwise noted on the product label, store the product under normal lab conditions in its tightly closed, original container. Do not pipet directly from the container or return unused portions to the container.

Instructions for Handling and Use - Please refer to the associated product label and Safety Data Sheet (SDS) for information regarding safety and handling of this product.

Preparation - All products are manufactured and tested according to established, documented procedures and methodology. Production documentation records manufacturing data, raw material traceability and testing history on a per lot basis. Balances, thermometers, and glassware are calibrated before first use and on a regular schedule with references traceable to NIST standards.

Submitted by: Greg Albright, Chemist Supervisor



An ISO9001:2008 certified company. Registration # 0306-01 01/30/2020 2:12 PM

Form #17.13 07/28/2016