



Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

12/15/2020 14:32:48 GMT
Page 1 of 1

Customer Name: Bel-Art Products-NJ
P.O.# MD61475

Processing Facility:

Salem

Work Order # 2791235
Sales Order # 2652572

23.5-50.0 kGy

BelArt, Gamma Treatment

Irradiation Date/Time:

12/14/2020 15:19:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
102 000	91	CA	369460000	STERILEWARE,SAMPLING SPOON,3 T	J346767	NA
103 000	1	CA	378470300	CYLINDER,PS,CLONING,STERILE,LA	J347255	NA
109 000	120	CA	369040000(20)	STERILEWARE,SCOOP,PS,4OZ	J346491	NA
	212	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	23.5 kGy	50.0 kGy	25.8 KGY	Pass	JROSARIO	12/14/2020 22:50:04 GMT
		Reason Code Test				JOSE ROSARIO	
450.00	Maximum Dose	23.5 kGy	50.0 kGy	44.1 KGY	Pass	JROSARIO	12/14/2020 22:50:53 GMT
		Reason Code Test				JOSE ROSARIO	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: AMY L. EDWARDS
Reason: Work Order Completions

Date: 12/15/2020 14:29:41 GMT