



Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

12/30/2020 13:23:50 GMT
Page 1 of 1

Customer Name:	Bel-Art Products-NJ	Processing Facility:	Salem	Work Order #	2798393
P.O.#	MD61620			Sales Order #	2661259
23.5-50.0 kGy		BelArt, Gamma Treatment	Irradiation Date/Time:	12/29/2020 12:13:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101 000	92	CA	369300000(30)	STERILEWARE,SAMPLING SPATULA,P	J346161	NA
102 000	140	CA	369060000(20)	STERILWARE,SCOOP,PS,8 OZ	J346964	NA
103 000	240	CA	369440000(30)	STERILEWARE,SAMPLING SPOON,1 T	J347215	NA
104 000	10	CA	369441010	STERILEWARE®,SMPLSPOON,1TSP	J347884	NA
105 000	20	CA	369021010	STERILEWARE®,SCOOP,PS,2OZ DBL	J347876	NA
117 000	40	CA	369020000(20)	STERILEWARE,SCOOP,PS,2OZ	J347873	NA
118 000	1	CA	378470100(1)	CYLINDER,PS,CLONING,STERILE,SM	J346815	NA
	543	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450 00	Minimum Dose	23.5 kGy	50.0 kGy	25.2 KGY	Pass	LSLATER	12/30/2020 00:56:55 GMT
		Reason Code Test				LAMAR SLATER	
450 00	Maximum Dose	23.5 kGy	50.0 kGy	44.6 KGY	Pass	LSLATER	12/30/2020 00:57:15 GMT
		Reason Code Test				LAMAR SLATER	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: AMY L. EDWARDS
Reason: Work Order Completions

Date: 12/30/2020 13:20:42 GMT