



Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
 TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

02/22/2021 13:58:33 GMT
 Page 1 of 1

Customer Name: Bel-Art Products-NJ Processing Facility: Salem Work Order # 2829281
 P.O.# MD62175 Sales Order # 2700055
 23.5-50.0 kGy BelArt, Gamma Treatment Irradiation Date/Time: 02/21/2021 15:30:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
102.000	210	CA	369460000	STERILEWARE,SAMPLING SPOON,3 T	J348354	NA
103.000	25	CA	136800070	Flowmi Cell Strainer	J347630	NA
111.000	280	CA	369020000(20)	STERILEWARE,SCOOP,PS,2OZ	J349125	NA
112.000	32	CA	136800040(24)	Flowmi Tip Strainers, 40 uM	J348102	NA
	547	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	Signed By User	Date /Time
450.00	Minimum Dose	23.5 kGy	50.0 kGy	26.0 KGY	Pass	LSLATER	02/21/2021 23:46:34 GMT
		Reason Code Test				LAMAR SLATER	
450.00	Maximum Dose	23.5 kGy	50.0 kGy	40.2 KGY	Pass	LSLATER	02/21/2021 23:46:57 GMT
		Reason Code Test				LAMAR SLATER	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: AMY L. EDWARDS
 Reason: Work Order Completions

Date: 02/22/2021 13:55:08 GMT