

Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

04/28/2021 16:45:13 GMT
Page 1 of 1

Customer Name: **Bel-Art Products-NJ** Processing Facility: **Salem** Work Order # **2869743**
P.O.# **MD62847** Sales Order # **2749991**
23.5-50.0 kGy BelArt, Gamma Treatment Irradiation Date/Time: **04/27/2021 07:07:00 GMT**

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
103.000	30	CA	367400005	STRLWR,ECON SMPL SPN,PS 20ML	J350023	NA
105.000	30	CA	367400004	STRLWR,ECON SMPE SPN,PS 8ML	J350021	NA
106.000	16	CA	136800070	Flowmi Cell Strainer	J350345	NA
108.000	70	CA	369040010	STERILEWARE,SCOOP,PS,4OZ,10/CA	J350655	NA
109.000	3	CA	378760000	VACCU-PETTE/96 PIPETTING DEVIC	J349810	NA
117.000	200	CA	369022002(20)	STERILWARE,SCOOP,RED,PS,2 OZ	J349801	NA
123.000	240	CA	369300000(30)	STERILEWARE,SAMPLING SPATULA,P	J349759	NA
124.000	20	CA	369020000(20)	STERILEWARE,SCOOP,PS,2OZ	J350582	NA
125.000	20	CA	369040000(20)	STERILEWARE,SCOOP,PS,4OZ	J350253	NA
126.000	26	CA	367400001(30)	STRLWR,ECON SMPL SPN,PS 1.25M	J351279	NA
127.000	30	CA	136800040(24)	Flowmi Tip Strainers, 40 uM	J350464	NA
	685	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	23.5 kGy	50.0 kGy	25.6 KGY	Pass	JNEWMAN	04/28/2021 08:31:55 GMT
450.00	Maximum Dose	Reason Code Test 23.5 kGy	50.0 kGy	43.8 KGY	Pass	Jeffrey Newman JNEWMAN	04/28/2021 08:32:08 GMT
		Reason Code Test				Jeffrey Newman	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: **AMY L. EDWARDS**
Reason: **Work Order Completions**

Date: **04/28/2021 16:44:48 GMT**

