

Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

06/01/2021 12:32:00 GMT
Page 1 of 1

Customer Name: Bel-Art Products-NJ Processing Facility: Salem Work Order # 2892923
P.O.# MD63186 Sales Order # 2778374
23.5-50.0 kGy BelArt, Gamma Treatment Irradiation Date/Time: 05/30/2021 09:57:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	140	CA	369440000(30)	STERILEWARE,SAMPLING SPOON,1 T	J350803	NA
102.000	160	CA	369022002(20)	STERILEWARE,SCOOP,RED,PS,2 OZ	J351598	NA
104.000	2	CA	369330002	STERILEWARE ECON SCOOP PS,83ML	J352405	NA
105.000	6	CA	367400003	STREWR,ECON SMPL SPN,PS 5ML	J352492	NA
106.000	16	CA	369440010(30)	STRLEWR,SMPL SPOON,1 TSP,PS,10/	J352502	NA
114.000	20	CA	369020006(20)	STERILEWARE@,METAL DETECT	J351919	NA
	344	CA	Total	SCOO		

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-Signed By -	Date /Time
450.00	Minimum Dose	23.5 kGy	50.0 kGy	26.8 KGY	Pass	SBARBIERI	05/31/2021 03:23:38 GMT
450.00	Maximum Dose	23.5 kGy	50.0 kGy	43.1 KGY	Pass	STEVEN BARBIERI	05/31/2021 03:23:48 GMT
		Reason Code Test				SBARBIERI	
		Reason Code Test				STEVEN BARBIERI	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: AMY L. EDWARDS
Reason: Work Order Completions

Date: 06/01/2021 12:31:45 GMT