

**Certificate of Processing**

STERIGENICS 75 Tilbury Road Salem NJ 08079  
TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

05/27/2021 20:46:30 GMT  
Page 1 of 1

Customer Name: **Bel-Art Products-NJ**  
P.O.# **MD63173**  
  
23.5-50.0 kGy

Processing Facility: **Salem**  
  
**BelArt, Gamma Treatment**  
Irradiation Date/Time:

Work Order # **2891064**  
Sales Order # **2776065**  
  
05/27/2021 14:12:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	70	CA	369460000	STERILEWARE,SAMPLING SPOON,3 T	J350373	NA
102.000	10	CA	369220000(30)	STERILEWARE,SCOOP SAMPLING SYS	J350983	NA
103.000	20	CA	367400310	STRLWR,ECON SMPL SPN,PS 5ML	J351917	NA
104.000	8	CA	367400110	STRLWR,ECON SMPL SPN,PS 1.25ML	J352256	NA
105.000	60	CA	369440000(30)	STERILEWARE,SAMPLING SPOON,1 T	J350803	NA
106.000	70	CA	369460010	Tablespoon 1.0	J352130	NA
107.000	40	CA	369022002(20)	STERILWARE,SCOOP,RED,PS,2 OZ	J351598	NA
108.000	115	CA	369022010	STERILEWARE,SCOOP,RED,PS,2OZ,1	J351463	NA
	393	CA	<b>Total</b>			

**Quality Test Summary**

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	23.5 kGy Reason Code Test	50.0 kGy	26.9 KGY	Pass	SBARBIERI	05/27/2021 19:09:21 GMT
450.00	Maximum Dose	23.5 kGy Reason Code Test	50.0 kGy	39.8 KGY	Pass	STEVEN BARBIERI SBARBIERI STEVEN BARBIERI	05/27/2021 19:09:31 GMT

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: **AMY L. EDWARDS**  
Reason: **Work Order Completions**

Date: 05/27/2021 20:46:11 GMT