



Certificate of Processing

STERIGENICS 75 Tilbury Road Salem NJ 08079
 TEL 856 935-1400 FAX 856 935-0542 www.sterigenics.com

R55480102

04/27/2023 17:28:41 GMT
 Page 1 of 1

Customer Name: Bel-Art Products-NJ
 P.O.# MD69444

Processing Facility: Salem

Work Order # 3333277
 Sales Order # 3295346

23.5-50.0 kGy

BelArt, Gamma Treatment

Irradiation Date/Time:

04/27/2023 11:12:00 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	20	CA	369060000(20)	STERILWARE,SCOOP,PS,8 OZ	J370416	NA
	Dose Map		125_10727-1	0C3 - 0E3x1.085		
102.000	60	CA	369060000(20)	STERILWARE,SCOOP,PS,8 OZ	J369730	NA
	Dose Map		125_10727-1	0C3 - 0E3x1.085		
103.000	30	CA	369440000(30)	STERILEWARE,SAMPLING SPOON,1 T	J363848	NA
	Dose Map		125_10746	0C3 - 0E3 x 1.08		
104.000	70	CA	369100000(30)	STERILEWARE STERILE SCOOP AND	J371222	NA
	Dose Map		125_10746	0C3 - 0E3 x 1.08		
105.000	189	CA	369400000(63)	STERILEWARE,SAMPLING SPOON,1/4	J368175	NA
	Dose Map		125_10744	0C3 - 0E3x1.082		
106.000	60	CA	367400004	STRLWR,ECON SMPE SPN,PS 8ML	J371635	NA
	Dose Map		125_10699-1	TE4-0C7		
107.000	6	CA	14660-1080	STERILEWARE PP POWDER FUNNEL,	J371497	NA
	Dose Map		125_10748	TE4 - 0C3		
108.000	11	CA	367400005	STRLWR,ECON SMPL SPN,PS 20ML	J372185	NA
	Dose Map		125_10699-1	TE4-0C7		
109.000	200	CA	369441010	STERILEWARE®,SMPLSPOON,1TSP	J370539	NA
	Dose Map		125_10877	0C3 - 0E3		
110.000	54	CA	369420010(B)	0.50 teaspoon	J370966	NA
	Dose Map		125_11179	0C3-0E4		
111.000	160	CA	369460010	Tablespoon 1.0	J370709	NA
	Dose Map		125_11368	**Triplicate Map Required**		
122.000	60	CA	369040006(20)	STERILEWARE®,METAL DETECT	J370414	NA
	Dose Map		125_10885	SCOO		
	920	CA	Total	0C3 - 0E3 x 1.14		

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	Signed By	Date /Time
450.00	Maximum Dose	23.5 kGy	50.0 kGy	40.9 KGY	Pass	SBARBIERI	04/27/2023 15:37:08 GMT
450.00	Minimum Dose	Reason Code Test 23.5 kGy	50.0 kGy	26.2 KGY	Pass	STEVEN BARBIERI	04/27/2023 15:37:25 GMT
		Reason Code Test				SBARBIERI	
						STEVEN BARBIERI	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: LESLIE GLESSNER
 Reason: Work Order Completions

Date: 04/27/2023 17:27:42 GMT