

Globe Scientific certifies that this product, made from resin material, has been manufactured under controlled conditions and is in compliance with all applicable procedures and specifications.

Item #: 135030

Lot #: K108070-03

Expiration Date: N/A

Product Descriptions: Transfer pipet, 7.0mL, 160mm large bulb, grad to 3mL, 500/bx


Component Material: Low Density Polyethylene (LDPE)

ISO Manufacturing ISO 9001:2015, ISO 13485:2016

Globe Scientific certifies that this product has been tested and is completely free of the following where applicable:

Statements:

- We hereby confirm that the LDPE used for this product is in conformity with RoHS and CONEG standards.
- We hereby confirm that the LDPE used for this product does not contain Phthalates or Estrogenic Chemicals, (such as BPA).
- We hereby confirm that the LDPE used for this product does not contain any Trace Metals.
- We hereby confirm that the LDPE used for this product does not contain Conflict Minerals as identified in 1502 of the Dodd-Frank Wall Street Reform and Consumer Protection Act.
- This product does not contain any microorganisms, viruses or prions.
- This product does not contain any antigens derived from, or antibodies raised against, microorganisms, viruses or prions.
- This product does not include (and is not accompanied by) any additional external reagents containing animal or microbial derived material.
- The resin used complies with US FDA 21 CFR 177.1520.
- Conflict Mineral Compliance - Cassiterite (for tin), wolframite (for tungsten), coltan (for tantalum), and gold ore from conflict regions are not used in this product.
- Proposition 65 -This product does not contain the substances regulated by California Proposition 65 in concentrations resulting in exposures above the safe harbor levels determined by OEHHA. As such, this product does not require a California Proposition 65 warning label.



Doug Contreras
Director of Operations

12/29/21

Date





Certificate of Analysis

Globe Scientific certifies that this product, made from resin material, has been manufactured under controlled conditions and is in compliance with all applicable procedures and specifications.

Item #: 1515

Lot #: A1087463070707

Expiration Date: N/A

Product Descriptions: Culture tube, 16x125mm, 19mL borosilicate, 4x250/cs

Component Material: Borosilicate Glass

Manufacturing Standard(s): ISO 9001:2015

Globe Scientific certifies that this product has been tested and is completely free of the following where applicable:

Proposition 65:

The material does not contain any of the chemical substances listed by the State of California in Proposition 65.

Latex:

We confirm that this product is Latex Free.

Plasticizers:

We confirm that no plasticizers are used in the production of this product.

Doug Contreras
Director of Operations

12/29/21

Date



Globe Scientific certifies that this product, made from resin material, has been manufactured under controlled conditions and is in compliance with all applicable procedures and specifications.

Item #: 2801

Lot #: 21P2021101

Expiration Date: June 30, 2025

Product Descriptions: Loop, 1uL w/needle, flexible natural, STR, 20/peel pack

Component Material: Polypropylene (PP)

Standard: ISO 9001:2015

Sterilization Method / Range: Ethylene Oxide Treatment

Sterility Assurance Level (SAL): 10^{-6}


Globe Scientific certifies that this product has been tested and is completely free of the following where applicable:

Sterility:

Globe Scientific declares that the device listed above, has been terminally sterilized with validated ethylene oxide treatment in accordance with EN 556; EN ISO 11135 standards.

Calibration:

Globe Scientific certifies that the inoculation loops listed above have been manufactured and tested under tightly controlled conditions and confirms the delivery of the indicated volume $\pm 15\%$ in accordance with the highest standards of Good Manufacturing Practice.



Doug Contreras
Director of Operations

12/29/21

Date





Certificate of Analysis

Globe Scientific certifies that this product, made from resin material, has been manufactured under controlled conditions and is in compliance with all applicable procedures and specifications.

Item #: 2801

Lot #: 21P2021102

Expiration Date: August 31, 2025

Product Descriptions: Loop, 1uL w/needle, flexible natural, STR, 20/peel pack

Component Material: Polypropylene (PP)

Standard: ISO 9001:2015

Sterilization Method / Range: Ethylene Oxide Treatment

Sterility Assurance Level (SAL): 10^{-6}

Globe Scientific certifies that this product has been tested and is completely free of the following where applicable:

Sterility:

Globe Scientific declares that the device listed above, has been terminally sterilized with validated ethylene oxide treatment in accordance with EN 556; EN ISO 11135 standards.

Calibration:

Globe Scientific certifies that the inoculation loops listed above have been manufactured and tested under tightly controlled conditions and confirms the delivery of the indicated volume $\pm 15\%$ in accordance with the highest standards of Good Manufacturing Practice.

A handwritten signature in black ink, appearing to read "Doug Contreras", is written over a horizontal line.

Doug Contreras
Director of Operations

12/29/21

Date

