



Certificate of Analysis

Cetrimide Agar
Catalog Number: C03-107
Lot Number: 200127022301
Expiration: January 2023

3651 Clipper Mill Rd.
Baltimore, MD 21211
Phone (410) 467-9983
Fax (410) 467-5088

Physical And Chemical Tests

Appearance:	Homogeneous and free flowing
Color:	Beige
pH of a 4.53% solution at 25 °C: (7.2 +/- 0.2)	7.13

Microbiological Tests

<u>Microorganisms</u>	<u>Result*</u>
<i>Escherichia coli</i> ATCC 8739	Inhibited
<i>Pseudomonas aeruginosa</i> ATCC 10145	Yel-gr colonies
<i>Pseudomonas aeruginosa</i> ATCC 27853	Yel-gr colonies
<i>Pseudomonas aeruginosa</i> ATCC 9027	Yel-gr colonies
<i>Staphylococcus aureus</i> ATCC 25923	Inhibited

(Yel-gr = yellow-green)

* The following results were obtained from type cultures after incubation at a temperature of 32.5 +/-2.5 °C and observed after 18 hours.

Laboratory Results: Satisfactory

Country of Manufacture:	USA
Testing performed by:	E.Fox/H.McGinnis
Completed on:	2-10-20
QC Release by:	Elizabeth Fox